

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Sensus GmbH Hannover
Meineckestrasse 10
30880 Laatzen
Germany

Measuring instrument A model of a **water meter**

Type : MeiMag

Environment classes : M1 / E2

Temperature range liquid : +0,1 °C / +50 °C

Temperature range ambient : -25 °C / +55 °C

Measuring range : See § 1.2 of the description

Further properties are described in the annexes:
– Description T10334 revision 0;
– Documentation folder T10334-1.

Valid until 17 December 2020

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
17 December 2010


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.