

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Ningbo Water Meter Co. LTD. No. 99, Lane 268, Beihai Road Ningbo 315033 China
Measuring instrument	A water meter Type : Prepayment Water Meter Model : LXSZ-15 ; LXSZ-20 ; LXSZ-25 Destined for the measurement of : Water Measuring Range : See paragraph 1.2 of the description. Environment classes : M1 / E1 Temperature classes : T30 and T50 Temperature range ambient : +5 °C / +55 °C Further properties are described in the annexes: – Description T10525 revision 4; – Documentation folder T10525-2.
Valid until	30 October 2022
Remark	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
17 March 2014


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.