

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	ŞENSOYLAR Klape Tesisat Malzemeleri San. ve Tic. Ltd. Şti. Bakırcılar ve Prinçciler Sanayi Sitesi Menekşe Caddesi No:13 Beylikdüzü - Istanbul Turkey
Measuring instrument	A multi-jet water meter
Trade mark	: TÜRKOĞLU
Type	: MELEN xx yyy; ILGIN xx yyy
Measuring Range	: See §1.2 of the description
Environment classes	: M1 / E1
Temperature range water	: +0,1 °C / +50 °C (class T50) +0,1 °C / +90 °C (class T90)
Temperature range ambient	: -10 °C / +55 °C
Destined for the measurement of	: Water
	Further properties are described in the annexes: – Description T10642 revision 0; – Documentation folder T10642-1.
Valid until	11 July 2024

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
11 July 2014


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified
Body can be verified at [http://ec.europa.eu/
enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMi (see www.nmi.nl).

Reproduction of the complete
document only is permitted.