

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Emerson Process Management Flow B.V. Neonstraat 1 6718 WX Ede The Netherlands
Measuring instrument	<p>A Coriolis Gas Meter</p> <p>Brand : Micro Motion</p> <p>Type : CMFxxx (see paragraph 1.2 for the meaning of xxx), CNG050 and DS600, with MVD electronics (see paragraph 1.1 for details)</p> <p>Destined for the measurement of : fuel gases, with an actual density^[1] of 1,4 kg/m³ and higher, and supercritical ethylene with a density up to 450 kg/m³</p> <p>Minimum – Maximum flow rate (Q_{min} - Q_{max}) : see paragraph 1.2 of the description</p> <p>Accuracy class : Class 1,0</p> <p>Environment classes : M3 / E3</p> <p>Temperature range gas : -40 °C / +150 °C for CMF200/300/350/400/HC2/HC3/HC4/DS600; -10 °C / +50 °C for CMF025/050/100; -25 °C / +55 °C for CNG050.</p> <p>Ambient temperature range : -40 °C / +55 °C</p> <p>Further properties are described in the annexes: – Description T10020 revision 15; – Documentation folder T10020-11.</p>
Valid until	8 May 2027
Remark	This revision replaces the previous versions, except for the documentation folder.

[1] See chapter 1.2.1 for more information on the actual minimum density.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
22 March 2022

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.