

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Natural Gas Solutions North America, LLC 16240 Port Northwest Drive Houston, Texas 77041 United States of America
Measuring instrument	A Rotary Displacement Gas Meter Type : Roots® Meter, L2-MA or Series B4 Destined for the measurement of : Gas volume Accuracy class : 1,0 Environment classes : M1/ E1 Temperature range : -40 °C / +70 °C, or -25 °C / +70 °C (only G1600 meter)
	Further properties are described in the annexes: – Description T10035 revision 8; – Documentation folder T10035-4.
Valid until	24 April 2027
Remarks	This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
3 December 2018



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

Reproduction of the complete document only is permitted.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

