


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Pietro Fiorentini S.p.A. Via E. Fermi, 8/10 36057 – Arcugnano (VI) Italy
Measuring instrument	An electronic gas-volume conversion device (EVCD) , intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gasmeter. Type : FloWEB Manufacturer's mark or name : Pietro Fiorentini Conversion principle : PTZ Device type : 1 (complete system) or 2 (separate components) Ambient temperature range : -10 °C / +40 °C Designed for : Non-condensing humidity Environment classes : M2 / E2 The intended location for the instrument is closed. Further properties are described in the annexes: – Description T10113 revision 16; – Documentation folder T10113-15.
Valid until	26 August 2018
Remarks	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
14 September 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.