


Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Elster - Instromet B.V. Munstermanstraat 6 7064 KA Silvolde The Netherlands
Measuring instrument	A Turbine Gas Meter Type : SM-RI-2 / SM-RI-2X Destined for the measurement of : Gas volume Accuracy class : Class 1,0 Environment classes : M2 / E1 Ambient temperature range : -25 °C / +70 °C Gas temperature range : -25 °C / +70 °C Further properties are described in the annexes – Description T10119 revision 2 – Documentation folder T10119-3
Valid until	24 April 2018
Remarks	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
4 July 2011

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).

Reproduction of the complete document only is permitted.