



EU - type examination certificate

Number **T10120** revision 8
Project number 1902352
Page 1 of 1

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Aerzener Maschinenfabrik GmbH
Reherweg 28
D-31855 Aerzen
Germany

Measuring instrument **A Rotary Displacement Gas Meter**
Type : DKZ
Manufacturer's mark or name : Aerzener Maschinenfabrik GmbH
Accuracy class : Class 1,0 or Class 1,5
Environment classes : M2 / E2
Ambient temperature range : -10 °C / +40 °C
Gas temperature range : -10 °C / +40 °C
Destined for the measurement of : Gas volume

Further properties are described in the annexes:
– Description T10120 revision 8;
– Documentation folder T10120-3.

Valid until 16 June 2028

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

NMI Certin B.V., Notified Body number 0122
16 June 2018


C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

