

# EC-type examination certificate



Nederlands Meetinstituut

Number **T10135** Revision 2  
Project number: 810783  
Page 1 of 2

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body number 0122

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer Elektrometal S.A.  
ul. Stawowa 71  
43-400 Cieszyn  
Poland

In respect of **A diaphragm gas meter**

Manufacturer : Elektrometal S.A.  
Type : EM-G1,6; EM-G2,5 or EM-G4

Characteristics Destined for the measurement of : gas volume  
Accuracy class : Class 1,5  
Temperature range : -25°C / +55°C  
Mechanical environment class : M2  
Electromagnetic environment class : E1

G-value	$Q_{max}$ [m <sup>3</sup> /h]	$Q_{min}$ [m <sup>3</sup> /h]	$Q_t$ [m <sup>3</sup> /h]
4	6	0,040	0,6
2,5	4	0,025	0,4
1,6	2,5	0,016	0,25

In the description number T10135 Revision 2 further characteristics are described.

Valid until 6 June 2018

**NMI Certin BV**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
PO Box 394  
3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.

