

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Metrologiewet (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer iMeter
Markenweg 7
7051 HS Varsseveld
The Netherlands

In respect of A model of an **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gasmeter
Type : MEVC

Characteristics Conversion principle : PTZ
Device type : 1 (complete system)
Ambient temperature range : -25 °C/+55 °C
Designed for : condensing humidity
Mechanical environment class : M2
Electromagnetic environment class : E2
The intended location for the instrument is open.

In the description number T10137 revision 0 further characteristics are described.

Valid until 11 June 2019

Description and documentation The instrument is described in the description number T10137 revision 0 and documented in the documentation folder T10137-1, appertaining to this EC-type examination certificate.

The Notified Body no. 0122
NMI Certin, 11 June 2009

Ing. C. Oosterman
Head Certification Board