

EC-type examination certificate

Number T10150 Revision 1 Project number: 805513

Page 1 of 1

Issued by NMi Certin B.V.

> Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands

Notified Body number 0122

In accordance

with

The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive

2004/22/EC on measuring instruments (MID).

Manufacturer Kalekalip Makina ve Kalip Sanayi A.S.

Sefaköy, Tevfikbey Mah. Istiklal Cad. No:13

34295-K. Çekmece Istanbul

Turkey

In respect of A diaphragm gas meter

> Manufacturer Kalekalip Makina ve Kalip Sanayi A.S.

Type SI-G1,6, SI-G2,5 and SI-G4

Characteristics Destined for the measurement of

gas volume Accuracy class Class 1,5 Temperature range -25°C / +55°C

Mechanical environment class M2 Electromagnetic environment class : E1

G-value	Q _{max} [m ³ /h]	Q _{min} [m³/h]	Q _t [m³/h]
4	6	0,040	0,6
2,5	4	0,025	0,4
1,6	2,5	0,016	0,25

In the description number T10150 Revision 1 further characteristics are described.

Valid until

17 June 2018

Description and documentation

The gasmeter is described in the description number T10150 Revision 1 and documentated in the documentation folder T10150-1, appertaining to this

EC-type examination certificate.

Remark

This version replaces the former version, with the exception of the documentation folder.

Dordrecht, 26 February 2009

NMi Certin B.W

Ing. C. Oosterman

manager Product Certification

NMi Certin BV

Hugo de Grootplein 1 3314 EG Dordrecht PO Box 394 3300 AJ Dordrecht, NL T +31 78 6332332 F+31 78 6332309 certin@nmi nl www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi")

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party

Reproduction of the complete document is

