

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer SMF Industrial Products and Systems Inc.
1467 Cad. (Elektrokent) No 18, Ostim,
06105 Ankara
Turkey

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : RFLO VC211LL
Conversion principle : PTZ
Device type : 2 (separate transmitters)
Ambient temperature range : -25 °C / +55 °C
Designed for : non-condensing humidity
Environment classes : M2 / E2
The intended location for the instrument is closed.

Further properties are described in the annexes:
- Description T10169 revision 2;
- Documentation folder T10169-1.

Valid until 19 February 2019

Remarks This revision replaces the earlier versions, with the exception of the documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
22 April 2016



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see www.nmi.nl).

Reproduction of the complete document only is permitted.