

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Elster-Instromet B.V. Munstermanstraat 6 7064 KA Silvolde The Netherlands
Measuring instrument	A Rotary Displacement Gas Meter Type : IRM-xxxx, see paragraph 1.2.2 for a detailed explanation of xxxx. Destined for the measurement of : Gas volume Accuracy class : Class 1,0 or 1,5 Environment classes : M2/E2 Temperature range : -25 °C / +55 °C Further properties are described in the annexes: – Description T10199 revision 3; – Documentation folder T10199-3.
Valid until	11 June 2019
Remarks	This revision replaces the earlier versions, including the documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
31 August 2018



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

Reproduction of the complete document only is permitted.

The designation of NMI Certin BV.as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>