

- Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:
- Manufacturer SAMGAS S.r.l.
Strad Provinciale 33, I-20080 Vernate, Italy
- Measuring instrument **A Diaphragm Gas Meter**
- Type : RS / 2001 LA (steel housing)
RS / 2001 AL (aluminium housing)
RS / 2,4 (steel housing)
RSE / 2001 LA (steel housing)
RSE / 2,4 (steel housing)
- Destined for the measurement of : Gas volume
- Accuracy class : See description paragraph 1.2.1
- Environment classes : M2 / E1
- Location : Closed
- Temperature range : See description paragraph 1.2.1


G-value	Q _{max} [m ³ /h]	Q _{min} [m ³ /h]	Q _t [m ³ /h]	V [dm ³]
6	10	0,060	1,0	2,4
4	6	0,040	0,6	2,4
4	6	0,040	0,6	1,2
2,5	4	0,025	0,4	1,2
1,6	2,5	0,016	0,25	1,2

Further properties are described in the annexes:

- Description T10226 revision 5;
- Documentation folder T10226-5.

- Valid until 16 September 2019
- Remarks This revision replaces the earlier versions, including its documentation folder.

- Issuing Authority **NMI Certin B.V., Notified Body number 0122**
20 September 2013



C. Oosterman
Head Certification Board