

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of module B of Directive 2004/22/EC, to:

Manufacturer Destas Dijital Elektronik San. ve Tic. A.S.
Seyhli Mah. Mimarsinan Cad. No:33
34906 PENDIK ISTANBUL
Turkey

Measuring instrument **A Diaphragm Gas Meter**

Type	:	DESGAM4
Destined for the measurement of	:	Gas volume
To be used in	:	Closed locations
Accuracy class	:	Class 1,5
Environment classes	:	M2 / E1
Ambient temperature range	:	- 10 °C / +40 °C
Gas temperature range	:	- 10 °C / +40 °C

G-value	Q _{max} [m ³ /h]	Q _{min} [m ³ /h]	Q _t [m ³ /h]
4	6	0,040	0,6

Further properties are described in the annexes
– Description T10247 revision 0
– Documentation folder T10247-1

Valid until 26 January 2020

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
26 January 2010


C. Oosterman
Head of the Certification Board