

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Sagemcom Italia S.p.A.  
Località Garavelle, frazione Santa Lucia  
06012 Città di Castello (PG)  
Italy

Measuring instrument A **Diaphragm Gas Meter**  
Type : lpn / S80 AT or MGxx AT  
lpn / S80 AL or MGxx AL  
PhN 10 / 20 / 30 / 60 / 100  
MG40 / 65 / 100

Manufacturer's mark or name : Sagemcom Italia S.p.A.

Brand : Siconia

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M1 / E1

Temperature range : - 25 °C / +55 °C

Location : See description §1.2

Further properties are described in the annexes:

- Description T10258 revision 11;
- Documentation folder T10258-9.

Valid until 12 April 2030

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
12 November 2020

Certification Board