

**Issued by** NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 9 of Directive 2004/22/EC, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2004/22/EC, to:

**Manufacturer** Samgas s.r.l.  
Strada Provinciale 33 – Km 0,6  
20080 – Vernate (MI)  
Italy

**Measuring instrument** **A Diaphragm Gas Meter**

Type : RS/....

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Location : Closed

Temperature range : - 25 °C / +55 °C

G-value	Q <sub>max</sub> [m <sup>3</sup> /h]	Q <sub>min</sub> [m <sup>3</sup> /h]	Q <sub>t</sub> [m <sup>3</sup> /h]	V [dm <sup>3</sup> ]
100	160	1	16	120
65	100	0,65	10	60
40	65	0,4	6,5	35
25	40	0,25	4	20
16	25	0,16	2,5	10
10	16	0,1	1,6	10
4	6	0,04	0,6	1,2

Further properties are described in the annexes  
– Description T10294 revision 0  
– Documentation folder T10294-1

**Valid until** 29 June 2020

**Issuing Authority** **NMI Certin B.V., Notified Body number 0122**  
29 June 2010



G. Oosterman  
Head Certification Board