

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Rongcheng Yu Xiang Industrial Co., Ltd.
118 YueHu Road, Rongcheng city
The People's Republic of China

Measuring instrument **A Diaphragm Gas Meter**

Type : G2,5B

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Temperature range : - 25 °C / +55 °C

Location : open

G-value	Q_{\max} [m ³ /h]	Q_{\min} [m ³ /h]	Q_t [m ³ /h]	V [dm ³]
2,5	4	0,025	0,4	1,2

Further properties are described in the annexes:
– Description T10387 revision 0;
– Documentation folder T10387-1.

Valid until 18 October 2022

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
18 October 2012

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the applicant
shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.