

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Itron GmbH
Hardeckstrasse 2
D-76185 Karlsruhe
Germany

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be
used for gas volume conversion as a sub-assembly (according to article 4 of
the MID) of a gas meter.

Type : Corus Compact
Conversion principle : T, PT or PTZ
Device type : 1 (complete system)
Ambient temperature range : -25 °C / +55 °C
Designed for : condensing humidity
Environment classes : M2 / E2
The intended location for the instrument is open.

Further properties are described in the annexes:
– Description T10388 revision 2;
– Documentation folder T10388-3.

Valid until 25 October 2021

Remarks This revision replaces the earlier versions, including its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
25 October 2013



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.