

Issued by NMI Certin B.V.,  
 designated and notified by the Netherlands to perform tasks with respect to  
 conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
 having established that the Measuring instrument meets the applicable  
 requirements of Directive 2014/32/EU, to:

Manufacturer Zenner Gas S.r.l.  
 Via S. Eufemia, 276  
 66010 – Fara Filiorum Petri (CH)  
 Italy

Measuring instrument A **Diaphragm Gas Meter** with optionally electronic temperature  
 conversion

Type : BK 04S/R01 or BK 04S/R02

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Location : Closed

Temperature range : -25 °C / +55 °C

G-value	Q <sub>max</sub> [m <sup>3</sup> /h]	Q <sub>min</sub> [m <sup>3</sup> /h]	Q <sub>t</sub> [m <sup>3</sup> /h]	V [dm <sup>3</sup> ]
4	6	0,040	0,6	1,2
2,5	4	0,025	0,4	1,2
1,6	2,5	0,016	0,25	1,2

Option: mounted with a device for gas volume conversion.

Further properties are described in the annexes:

- Description T10391 revision 6;
- Documentation folder T10391-4.

Valid until 2 September 2021

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
 19 December 2017



C. Oosterman  
 Head Certification Board