

EC type-examination Certificate

Number **T10426** revision 0
Project number 11200344
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Flow Meter Group B.V. Meniststraat 5c 7091 ZZ Dinxperlo The Netherlands
Measuring instrument	A Turbine Gas Meter including an electronic indication with optional volume conversion Type : ETTM Destined for the measurement of : Gas volume Accuracy class : Class 1,0 Temperature range gas : See §1.2.1 of the description Temperature range ambient : See §1.2.1 of the description Environmental classes : M2 / E2 Further properties are described in the annexes: – Description T10426 revision 0; – Documentation folder T10426-1.
Valid until	20 January 2024

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
20 January 2014



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMi (see www.nmi.nl).

Reproduction of the complete
document only is permitted.