

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Emerson Process Management
Daniel Measurement and Control Inc.
11100 Brittmoore Park Drive
Houston, TX 77041
United States of America

Measuring instrument An **Ultrasonic Gas Meter**
Type : model 3400 / 3422, mark III
Destined for the measurement of : Gas volume
 $Q_{\min} - Q_{\max}$: see paragraph 1.2 of the
description
Accuracy class : Class 1,0
Environment classes : M2 / E2
Temperature range gas : -25°C / +55°C
Temperature range ambient : -25°C / +55°C
Designed for : Condensing humidity
Intended location : Open

Further properties are described in the annexes
– Description T10490 revision 0
– Documentation folder T10490-1

Valid until 17 December 2022

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
17 December 2012

C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.