

# EC type-examination Certificate

Number **T10534** revision 0  
Project number 12200811  
Page 1 of 1

|                      |   |      |  |
|----------------------|---|------|--|
| Issued by            | NMI Certin B.V.,<br>designated and notified by the Netherlands to perform tasks with respect to<br>conformity modules mentioned in article 9 of Directive 2004/22/EC, after<br>having established that the Measuring instrument meets the applicable<br>requirements of Directive 2004/22/EC, to: |      |  |
| Manufacturer         | Itron GmbH<br>Hardeckstrasse 2<br>76185 Karlsruhe<br>Germany  |      |  |
| Measuring instrument | <b>A Diaphragm Gas Meter</b>  | Type | : FGCI<br>(see also §1.2.1 of the description) |
|                      | Destined for the measurement of   | :    | Gas volume                                     |
|                      | Accuracy class  | :    | Class 1,5                                      |
|                      | Environment classes   | :    | M2 / E2  |
|                      | Temperature range   | :    | - 25 °C / +55 °C                               |
|                      | Location  | :    | Open   |
|                      | Further properties are described in the annexes:<br>– Description T10534 revision 0;<br>– Documentation folder T10534-1.  |      |  |
| Valid until          | 12 July 2023  |      |  |
| Issuing Authority    | <b>NMI Certin B.V., Notified Body number 0122</b><br>12 July 2013   |      |  |

C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that  
no liability is accepted and that the applicant  
shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [http://  
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge  
objection against this decision,  
within six weeks after the date of  
submission, to the general manager  
of NMI (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete  
document only is permitted.