

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Zenner Gas S.r.l.
Via Aterno, 122
66020 San Giovanni Teatino (CH)
Italy

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : PULSAR CK 01A/R01
Conversion principle : PT
Device type : 1 (complete system)
Ambient temperature range : -25 °C / +55 °C
Designed for : non-condensing humidity
Environment classes : M2 / E2
The intended location for the instrument is open.

Further properties are described in the annexes:
– Description T10552 revision 5;
– Documentation folder T10552-3.

Valid until 21 January 2024

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
9 July 2019


C. Oosterman
Head Certification Board