

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer GFO Flow Solutions B.V.  
Magnesiumstraat 14  
6031 RV Nederweert  
The Netherlands

Measuring instrument **A Rotary Displacement Gas Meter**

Type : GFO-RM  
Destined for the measurement of : Gas volume  
Accuracy class : Class 1,0  
Environment classes : M2 / E1  
Temperature range : -25 °C / +55 °C  
Maximum pressure : 20 bar(g)

Further properties are described in the annexes:  
- Description T10620 revision 2;  
- Documentation folder T10620-2.

Valid until 4 March 2024

Remark This revision replaces the earlier versions, including its documentation  
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
19 October 2017



C. Oosterman  
Head Certification Board