

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 9 of Directive 2004/22/EC, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2004/22/EC, to:

Manufacturer CPL Concordia Soc. Coop.  
Via A. Grandi, 39  
41033 Concordia sulla Secchia (MO)  
Italy

Measuring instrument **A Diaphragm Gas Meter**

Type : EG10 or EG10 SMART

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Temperature range : See description §1.2.1

Location : Closed

G-value	$Q_{\max}$ [m <sup>3</sup> /h]	$Q_{\min}$ [m <sup>3</sup> /h]	$Q_t$ [m <sup>3</sup> /h]	V [dm <sup>3</sup> ]	$p_{\max}$ [bar]
10	16	0,10	1,6	6	0,5

Further properties are described in the annexes:

- Description T10622 revision 1;
- Documentation folder T10622-2.

Valid until 30 January 2024

Remarks This revision replaces the earlier version, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
4 November 2014



C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that  
no liability is accepted and that the applicant  
shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [http://  
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge  
objection against this decision,  
within six weeks after the date of  
submission, to the general manager  
of NMI (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete  
document only is permitted.