

EC type-examination Certificate

Number **T10778** revision 0
Project number SO16200178
Page 1 of 1

Issued by	NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:		
Manufacturer	Tancy Instrument Group Co.,Ltd. No.3468, Tongfu Rd. Lingxi town, Cangnan county Wenzhou city, Zhejiang province China		
Measuring instrument	A turbine Gas Meter		
	Type	:	TBQJ
	Manufacturer's mark or name	:	Tancy Instrument Group Co.,Ltd.
	Destined for the measurement of	:	Gas volume
	Accuracy class	:	Class 1,0
	Environment classes	:	M1 / E1
	Gas temperature range	:	+5 °C / +55 °C
Ambient temperature range	:	+5 °C / +55 °C	
	Further properties are described in the annexes: – Description T10778 revision 0; – Documentation folder T10778-1.		
Valid until	19 April 2026		

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
19 April 2016



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMI (see www.nmi.nl).

Reproduction of the complete
document only is permitted.