

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Gas Souzan Ind. & Mfg. Co. Industrial Zone, Najafabad, Isfahan, Iran
Measuring instrument	An Ultrasonic Gas Meter
Type	: UIM-4F
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,0
Environment classes	: M1 / E2
Ambient temperature range	: -25 °C / +55 °C
Designed for	: Condensing humidity
Intended location	: Open
	Further properties are described in the annexes: – Description T10941 revision 0; – Documentation folder T10941-2.
Valid until	19 December 2026

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
19 December 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Reproduction of the complete document only is permitted.