

EU-type examination certificate

Number **T 10964** revision 2
Project number 3207569
Page 1 of 1

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
assessment modules mentioned in article 17 of Directive 2014/32/EU, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer MI-GAS S.r.l.
Via XX Settembre, 9
20080 ZIBIDO SAN GIACOMO (MI)
Italy

Measuring instrument **A Diaphragm Gas Meter**

Type : NPA12

Manufacturer's mark or name : MI-GAS S.r.l.

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Temperature range : - 25 °C / +55 °C

Location : Closed

G-value	Q _{max} [m ³ /h]	Q _{min} [m ³ /h]	Q _t [m ³ /h]	V [dm ³]
4	6	0,040	0,6	1,2
2,5	4	0,025	0,4	1,2
1,6	2,5	0,016	0,25	1,2

Further properties are described in the annexes:
- Description T10964 revision 2;
- Documentation folder T10964-3.

Valid until 3 January 2027

Remarks
- This revision replaces the previous revisions;
- The documentation folder has been revised.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
9 January 2023

Certification Board