

|                      |  |
|----------------------|--|
| Issued by            | NMi Certin B.V.,<br>designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:  |
| Manufacturer         | EDMI Electronics Sdn Bhd<br>PL0226A, Jalan.<br>Cyber 1A, Senai Industrial Park Phase 3,<br>81400, Senai Johor,<br>Malaysia   |
| Measuring instrument | An <b>Ultrasonic Gas Meter</b><br>Type : GS-60A<br>Destined for the measurement of : Gas volume of natural gas,<br>type H, L, E<br>Accuracy class : Class 1,5<br>Environment classes : M1 / E2<br>Gas temperature range : - 10 °C / +55 °C<br>Ambient temperature range : - 10 °C / +55 °C<br>Designed for : Condensing humidity<br>Intended location : Open<br><br>Further properties are described in the annexes:<br>– Description T11110 revision 0;<br>– Documentation folder T11110-1. |
| Valid until          | 19 June 2027   |

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
19 June 2017



C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/hando/>

Reproduction of the complete document only is permitted.