

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Elster GmbH
D-55252 Mainz-Kastel
Steinernstr. 19-21
Germany

Measuring instrument **A Rotary Displacement Gas Meter**

Type	: IRM-3 DUO
Manufacturer's mark or name	: Elster
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,0 or 1,5
Environment classes	: M1/ E2
Temperature range	: -25 °C / +70 °C

Further properties are described in the annexes:
– Description T11553 revision 0;
– Documentation folder T11553-1.

Valid until 29 July 2029

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
29 July 2019


C. Oosterman
Head Certification Board