

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Sensia LLC  
7000 Nix Drive  
Duncan, Oklahoma,  
United States of America

Measuring instrument **Electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : QRATE Scanner 3100 &  
QRATE Scanner 3200

Manufacturer's mark or name : Sensia

Conversion principle : PTZ

Ambient temperature range : -40 °C / +70 °C

Designed for : Condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is open.

Further properties are described in the annexes

- Description T11867 revision 0
- Documentation folder T11867-1

Valid until 30 July 2031

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
30 July 2021

## Certification Board

**NMI Certin B.V.**  
Thijssesweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.