

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Daesung Measuring CO., LTD.
5 Taebong 1-gil, Hogye-myeon
Mungyeong-si, Gyeongsangbuk-do 36931
South Korea

Measuring instrument **Diaphragm gas meter**
Type : DDM-A-G1.6
Manufacturer's mark or name : Daesung
Destined for the measurement of : Gas volume
Accuracy class : Class 1,5
Environment classes : M1 / E1
Temperature range : - 10 °C / +55 °C
Location : Closed

Further properties are described in the annexes:

- Description T11920 revision 2;
- Documentation folder T11920-2.

Valid until 7 August 2030

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
8 December 2022

Certification Board