

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Landis+Gyr AG
Alte Steinhauserstrasse 18
CH-6330 Cham
Switzerland

Measuring instrument **Ultrasonic gas meter**
Type : E06*2* 56
E06*4* 56

Manufacturer's mark or name : Landis + Gyr

Destined for the measurement of : Gas volume of natural gas,
See description §1.2.2

Accuracy class : Class 1,5

Environment classes : M1 / E1

Temperature range : - 10 °C / +40 °C

Designed for : Condensing humidity

Location : Closed

Further properties are described in the annexes:

- Description T12158 revision 5;
- Documentation folder T12158-3.

Valid until 29 November 2031

Initially issued 29 November 2021

Remarks This revision replaces the previous version(s), including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
2 November 2023

Certification Board