


Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer EDM I Limited
47 Yishun Industrial Park A
Singapore 768724

Measuring instrument **Ultrasonic gas meter**

Type	: GC-25AU
Manufacturer's mark or name	:  EDM I
Destined for the measurement of	: Gas volume of natural gas, see description §1.2.2
Accuracy class	: Class 1,5
Environment classes	: M1 / E2
Temperature range	: - 10 °C / +55 °C
Designed for	: Condensing humidity
Location	: Open

Further properties are described in the annexes:
– Description T12198 revision 1;
– Documentation folder T12198-2.

Valid until 7 June 2032

Initially issued 7 June 2022

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
15 December 2023

Certification Board