


Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer gAilar
Kamerlingh Onnesweg 63
3316 GK Dordrecht
The Netherlands

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : gFlow 1500

Manufacturer's mark or name : 
always in control

Conversion principle : T, PT or PTZ

Ambient temperature range : -25 °C / +55 °C

Designed for : Condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is open.

- Further properties are described in the annexes:
- Description T12271 revision 1;
 - Documentation folder T12271-1.

Valid until 10 August 2033

Initially issued 10 August 2023

Remark This revision replaces the earlier versions, excluding its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
9 February 2024

Certification Board