

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer OMNI Flow Computers, Inc.
12320 Cardinal Meadow, Suite 180
Sugar Land, Texas, 77478-6218
United States of America

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : OMNI 4000 and OMNI 7000

Manufacturer's mark or name : OMNI Flow Computers, Inc.

Conversion principle : PTZ

Ambient temperature range : -10 °C / +55 °C

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is closed.

- Further properties are described in the annexes
 - Description T12338 revision 0
 - Documentation folder T12338-1

Valid until 27 December 2032

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
27 December 2022

Certification Board

NMI Certin B.V.,
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.