

## EU-type examination certificate



Number **T10203** revision 26 Project number 3469548 Page 1 of 1

: Flow-X/M, Flow-X/P, Flow-X/S, Flow-X/R,

Issued by

NMi Certin B.V.

+

designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

ABB B.V.

Achtseweg Zuid 151A 5651 GW Eindhoven The Netherlands

Measuring Instrument

An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a

gas meter.

Type

and Flow X/K : ABB B.V.

Manufacturer's name : ABB Conversion principle : PTZ

Ambient temperature range : see § 1.1 of the description
Designed for : non-condensing humidity

Mechanical environment class : M2 Electromagnetic environment class : E2

The intended location for the instrument is a closed one.

+

Further properties are described in the annexes:

- Description number T10203 revision 26:
- Documentation folder number T10203-14.

Valid until 11 February 2030

Initially issued 11 February 2010

Remark This revision replaces the previous versions, except for its documentation folder.



## **Issuing Authority**

www.nmi.nl

## NMi Certin B.V., Notified Body number 0122

23 December 2024



## **Certification Board**

NMi Certin B.V. This doc Thijsseweg 11 no liabi 2629 JA Delft manufa The Netherlands liability T +31 88 636 2332

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <a href="http://ec.europa.eu/growth/tools-databases/nando/">http://ec.europa.eu/growth/tools-databases/nando/</a>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





