

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Hangzhou Beta Meter Co., Ltd.
No.188 Songyuan Street
Lin'an District, Hangzhou
P.R. China

Measuring instrument **Diaphragm gas meter**

Type : G4, G2.5 and G1.6
G2.5 and G1.6 prepaid wireless

Manufacturer's mark or name : beta

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M1 / E2

Temperature range : - 10 °C / +55 °C

Designed for : Condensing humidity

Location : Closed

Further properties are described in the annexes:
– Description T10438 revision 7;
– Documentation folder T10438-5.

Valid until 16 April 2032

Initially issued 16 April 2012

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
19 March 2024

Certification Board