

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 17 of Directive 2014/32/EU, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer

Tancy Instrument Group Co., Ltd.
No. 198, Hualian Road, Cangnan Industrial Zone
Wenzhou City, Zhejiang Province
P.R. China

Measuring instrument

A Turbine Gas Meter

Type	:	TBQM
Manufacturer's mark or name	:	Tancy
Destined for the measurement of	:	Gas volume
Accuracy class	:	Class 1,0
Environment classes	:	M1 / E1
Gas temperature range	:	-10 °C / +55 °C
Ambient temperature range	:	-10 °C / +55 °C

Further properties are described in the annexes:

- Description T10488 revision 4;
- Documentation folder T10488-2.

Valid until

26 September 2024

Initially issued

26 September 2014

Remark

This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

NMi Certin B.V.
14 May 2024

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as
Notified Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed and
sealed. The digital signature can be
verified in the blue ribbon at the top
of the electronic version of this
certificate.