

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Tancy Instrument Group Co., Ltd.  
No. 198, Hualian Road, Cangnan Industrial Zone  
Wenzhou City, Zhejiang Province  
P.R. China

Measuring instrument **A Turbine Gas Meter**

Type	: TBQM
Manufacturer's mark or name	: Tancy
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,0
Environment classes	: M1 / E1
Gas and ambient temperature range	: +5 °C / +55 °C or -10 °C / +55 °C
	: See also section 1.1.6

Further properties are described in the annexes  
– Description T10488 revision 5  
– Documentation folder T10488-3

Valid until 26 September 2034

Initially issued 26 September 2014

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
17 February 2025

Certification Board