

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer ZENNER Metering Technology (Shanghai) Ltd.  
No. 800 Songda Road  
Qingpu District  
Shanghai  
P.R. China

Measuring instrument **Diaphragm gas meter**  
Type

: Atmos xxS  
(xx is G1.6, G2.5, G4, G4L, G6M,  
WG2.5 or WG6M)

Atmos HP xxA  
(xx is G1.6, G2.5, G4, G4L, G6M,  
WG2.5 or WG6M)

Manufacturer's mark or name : ZENNER

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M1 / E1

Temperature range : - 25 °C / +55°C

Location : Closed/Open

Further properties are described in the annexes

- Description T10551 revision 8;
- Documentation folder T10551-9.

Valid until 19 October 2025

Initially issued 19 October 2015

Remarks This revision replaces the earlier version(s), including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
29 August 2024

## Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.