

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Flonidan A/S  
Islandsvej 29  
8700 Horsens  
Denmark

Measuring instrument **Ultrasonic gas meter**  
Type

: SciFlo G4U  
SciFlo G6U  
SciFlo G4-6U

Manufacturer's mark or name : Flonidan A/S

Destined for the measurement of : Gas volume of natural gas,  
type H, L and E

Accuracy class : Class 1,5

Environment classes : M1 / E2

Gas temperature range : - 25 °C / +55 °C

Ambient temperature range : - 25 °C / +55 °C

Designed for : Condensing humidity

Location : Closed

Further properties are described in the annexes:

- Description T10801 revision 11;
- Documentation folder T10801-10.

Valid until 18 May 2028

Initially issued 18 May 2018

Remark This revision replaces the earlier versions, excluding its documentation  
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
28 August 2024

## Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that  
no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [http://  
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

This document is digitally signed  
and sealed. The digital signature  
can be verified in the blue ribbon at  
the top of the electronic version of  
this certificate.