

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Micro Motion, Inc.  
7070 Winchester Circle  
Boulder, CO 80301  
United States of America

Measuring instrument An **Ultrasonic Gas Meter**  
Type : 3414 / 3415 / 3416 / 3417 / 3418  
GUSM or Senior Sonic  
Manufacturer's mark : Rosemount  
Destined for the measurement of : Gas volume  
Accuracy class : Class 1.0  
Environment classes : M1 / E2  
Designed for : Condensing humidity  
Intended location : Open

Further properties are described in the annexes:

- Description T11159 revision 15.
- Documentation folder T11159-6.

Valid until 20 July 2027

Initially issued 20 July 2017

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
6 March 2024

## Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.