

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Secure Meters Ltd.
Pratapnagar Industrial Area
Udaipur 313003
India

Measuring instrument **Thermal-mass flow gas meter**
Type : Liberty Gas 120
Liberty Gas 200
Liberty Gas 220

Manufacturer's mark or name : **SECURE**

Destined for the measurement of : Gas volume of natural gas,
See description §1.2.2

Accuracy class : Class 1,5

Environment classes : M1 / E2

Temperature range : See description §1.2.1

Designed for : Condensing humidity

Location : See description §1.4.1

Further properties are described in the annexes

– Description T11223 revision 15

– Documentation folder T11223-10

Valid until 11 October 2028

Initially issued 11 October 2018

Remarks This revision replaces the earlier version(s), including its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
6 February 2025

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.