

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Secure Meters Ltd.
Pratapnagar Industrial Area
Udaipur 313003
India

Measuring instrument **Thermal-mass flow gas meter**
Type : Liberty Gas 120
Liberty Gas 200
Liberty Gas 220

Manufacturer's mark or name : 

Destined for the measurement of : Gas volume of natural gas,
See description §1.2.2

Accuracy class : Class 1,5

Environment classes : M1 / E2

Temperature range : See description §1.2.1

Designed for : Condensing humidity

Location : See description §1.4.1

Further properties are described in the annexes:

- Description T11223 revision 17;
- Documentation folder T11223-10.

Valid until 11 October 2028

Initially issued 11 October 2018

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
9 April 2026

Certification Board