

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer SICK Engineering GmbH
Bergener Ring 27
01458 Ottendorf-Okrilla
Germany

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : Flow-X/C

Conversion principle : PTZ

Manufacturer's mark or name : SICK

Ambient temperature range : see § 1.1 of the description

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description T11449 revision 11;
- Documentation folder number T11449-7.

Valid until 24 December 2028

Initially issued 24 December 2018

Remark This revision replaces the previous versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
7 May 2024

Certification Board