

EU-type examination certificate

Number **T11920** revision 3
Project number 3812476
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Daesung Measuring CO., LTD.
5 Taebong 1-gil, Hogye-myeon
Mungyeong-si, Gyeongsangbuk-do 36931
South Korea

Measuring instrument

Diaphragm gas meter

Type : DDM-A-G1.6
Manufacturer's mark or name : Daesung
Destined for the measurement of : Gas volume
Accuracy class : Class 1,5
Environmental classes : M1 / E1
Temperature range : -10 °C / +55 °C
Designed for : Condensing humidity
Location : Closed

Further properties are described in the annexes:

- Description T11920 revision 3;
- Documentation folder T11920-3.

Valid until

7 August 2030

Initially issued

7 August 2020

Remark

This revision replaces the earlier versions, including its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
26 May 2025

Certification Board