

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Euromisure WIKA Instruments S.a.s. di WIKA Italia S.r.l.  
Via G. Borghisani, 4  
26035 Pieve S. Giacomo (CR)  
Italy

Measuring instrument An **Ultrasonic Gas Meter**  
Type : FLC-UFL-4F  
Manufacturer's mark or name : Euromisure WIKA Instruments S.a.s. di WIKA Italia S.r.l.

Destined for the measurement of : Gas volume  
Accuracy class : Class 1,0  
Environment classes : M1 / E2  
Ambient temperature range : -25 °C / +55 °C  
Designed for : Condensing humidity  
Intended location : Open

Further properties are described in the annexes:  
– Description T11972 revision 3;  
– Documentation folder T11972-3.

Valid until 6 October 2030

Initially issued 6 October 2020

Remark This revision replaces the previous versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
1 November 2024

Certification Board