


Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer EDMi Limited  
47 Yishun Industrial Park A  
768724 SINGAPORE

Authorized representative EDMi Electronics Sdn Bhd  
PL0226A, Jalan Cyber 1A, Senai Industrial Park Phase 3  
81400 Senai Johor  
Malaysia

Measuring instrument **Ultrasonic gas meter**

Type : GR-8BU

Manufacturer's mark or name :  EDMi

Destined for the measurement of : Gas flow of natural gas. See  
description §1.2.2

Accuracy class : Class 1,5

Environment classes : M1 / E2

Gas temperature range : - 10 °C / +55 °C

Ambient temperature range : - 10 °C / +55 °C

Designed for : Condensing humidity

Further properties are described in the annexes  
– Description T12077 revision 5  
– Documentation folder T12077-1

Valid until 6 May 2031

Initially issued 6 May 2021

Remarks This revision replaces the earlier version(s), except for its documentation  
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
17 September 2024

Certification Board